## Message

From: Khan, Annie (DPH) [/O=COMMONWEALTH OF MASSACHUSETTS/OU=DPH/CN=RECIPIENTS/CN=AKHAN]

**Sent**: 10/21/2005 3:42:54 PM

To: 'Anthony'

Subject: RE: Perkin Elmer FTIR: Serial # 73816

Tony, thank you very much. It works fine.

Annie Khan SLI-DPH Drug Lab 305 South Street, Jamaica Plain, MA 02130 617-983-6631

From: Anthony [mailto:

Sent: Thursday, October 20, 2005 6:04 PM

To: 'Khan, Annie (DPH)'

Subject: RE: Perkin Elmer FTIR: Serial # 73816

Annie,

Turn the **instrument** and the computer off. Count to 5 then turn the instrument and the computer on again. It should be fine.

The power switch for the instrument is in the back right hand corner.

If you have problems call me Friday between 8Am to 10 AM at 508-234-0916.

Anthony Rosselli 508-234-0916 PerkinElmer